



ALS Technichem (HK) Pty Ltd
11/F., Chung Shun Knitting Centre,
1 - 3 Wing Yip Street,
Kwai Chung, N.T., Hong Kong
T: +852 2610 1044
F: +852 2610 2021
www.alsglobal.com

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION

CONTACT: MR K.W.FAN
CLIENT: ENVIROTECH SERVICES CO.
ADDRESS: RM 712, 7/F, MY LOFT
9 HOI WING ROAD,
TUEN MUN, N.T. HK

WORK ORDER: HK2550021
SUB-BATCH: 0
LABORATORY: HONG KONG
DATE RECEIVED: 18-Nov-2025
DATE OF ISSUE: 25-Nov-2025

GENERAL COMMENTS

The performance of the equipment stated in this report is checked with independent reference material and results compared against a calibrated secondary source.

The "Tolerance Limit" quoted is the acceptance criteria applicable for similar equipment used by the laboratory or quoted from relevant international standards.

The "Next Calibration Date" is recommended according to best practice principle as practised by the laboratory or quoted from relevant international standards.

The validity of equipment/ meter performance only applies to the result(s) stated in the report.

This report supersedes any previous report(s) with same work order number.

EQUIPMENT INFORMATION

Equipment information (Brand name, Model No., Serial No. and Equipment No.) is provided by client.
Equipment Type: Multifunctional Meter
Service Nature: Performance Check
Scope: Conductivity, Dissolved Oxygen, pH Value, Turbidity, Salinity and Temperature
Brand Name/ Model No.: [HORIBA]/ [U-53]
Serial No./ Equipment No.: [7D61KMPX]/ [N/A]
Date of Calibration: 25-November-2025

May

Ms. Cheng Sin Ying, May
Senior Chemist - Inorganics

This report shall not be reproduced except in full without the written approval of the laboratory.

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK2550021

SUB-BATCH: 0

DATE OF ISSUE: 25-Nov-2025

CLIENT: ENVIROTECH SERVICES CO.

Equipment Type: Multifunctional Meter

Brand Name/ Model No.: [HORIBA]/ [U-53]

Serial No./ Equipment No.: [7D61KMPX]/ [N/A]

Date of Calibration: 25-November-2025

Date of Next Calibration:

25-February-2026

PARAMETERS:

Conductivity

Method Ref: APHA (23rd edition), 2510B

Expected Reading ($\mu\text{S}/\text{cm}$)	Displayed Reading ($\mu\text{S}/\text{cm}$)	Tolerance (%)
146.9	153	+4.2
6667	6550	-1.8
12890	13300	+3.2
58670	60100	+2.4
Tolerance Limit (%)		± 10.0

Dissolved Oxygen

Method Ref: APHA (23rd edition), 4500O: G

Expected Reading (mg/L)	Displayed Reading (mg/L)	Tolerance (mg/L)
2.94	2.92	-0.02
5.18	5.15	-0.03
6.64	6.51	-0.13
Tolerance Limit (mg/L)		± 0.20

pH Value

Method Ref: APHA (23rd edition), 4500H: B

Expected Reading (pH unit)	Displayed Reading (pH unit)	Tolerance (pH unit)
4.0	3.93	-0.07
7.0	6.93	-0.07
10.0	9.93	-0.07
Tolerance Limit (pH unit)		± 0.20

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

May

Ms. Cheng Sin Ying, May
Senior Chemist - Inorganics

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK2550021

0

25-Nov-2025

ENVIROTECH SERVICES CO.

Equipment Type: Multifunctional Meter

Brand Name/

Model No.:

Serial No./

Equipment No.: [7D61KMPX]/ [N/A]

Date of Calibration: 25-November-2025

Date of Next Calibration:

25-February-2026

PARAMETERS:

Turbidity

Method Ref: APHA (23rd edition), 2130B

Expected Reading (NTU)	Displayed Reading (NTU)	Tolerance (%)
0	0.00	--
4	3.96	-1.0
40	43.2	+8.0
80	85.0	+6.3
400	429	+7.3
800	867	+8.4
Tolerance Limit (%)		±10.0

Salinity

Method Ref: APHA (23rd edition), 2520B

Expected Reading (ppt)	Displayed Reading (ppt)	Tolerance (%)
0	0.00	--
10	9.47	-5.3
20	19.48	-2.6
30	28.48	-5.1
Tolerance Limit (%)		±10.0

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

May

Ms. Cheng Sin Ying, May
Senior Chemist - Inorganics

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK2550021

SUB-BATCH: 0

DATE OF ISSUE: 25-Nov-2025

CLIENT: ENVIROTECH SERVICES CO.

Equipment Type: Multifunctional Meter

Brand Name/ Model No.: [HORIBA]/ [U-53]

Serial No./ Equipment No.: [7D61KMPX]/ [N/A]

Date of Calibration: 25-November-2025

Date of Next Calibration:

25-February-2026

PARAMETERS:

Temperature

Method Ref: Section 6 of International Accreditation New Zealand Technical Guide No. 3 Second edition March 2008: Working Thermometer Calibration Procedure.

Expected Reading (°C)	Displayed Reading (°C)	Tolerance (°C)
12.0	12.86	+0.9
22.5	21.50	-1.0
38.0	38.07	+0.1
Tolerance Limit (°C)		±2.0

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Ms. Cheng Sin Ying, May
Senior Chemist - Inorganics